

## Analytical Method Information

Analyte	MDL	Reporting Limit	Surrogate %R	Duplicate RPD	Matrix Spike %R	RPD	Blank Spike / LCS %R	RPD
<b>PCB Congener Screen in Water (GCMS SIM)</b>								
<b>Preservation:</b> <6°C, Na2S2O3 (If Chlorinated)								
<b>Container:</b> 1 L Amber Glass - 625/8270			<b>Amount Required:</b> 2000 ml			<b>Hold Time:</b> 14 days		
PCB - 8		10 ng/l						
PCB - 18		10 ng/l						
PCB - 28		10 ng/l						
PCB - 44		10 ng/l						
PCB - 52		10 ng/l						
PCB - 66		10 ng/l						
PCB - 77		10 ng/l						
PCB - 81		10 ng/l						
PCB - 101		10 ng/l						
PCB - 105		10 ng/l						
PCB - 114		10 ng/l						
PCB - 118		10 ng/l						
PCB - 123		10 ng/l						
PCB - 126		10 ng/l						
PCB - 128		10 ng/l						
PCB - 138		10 ng/l						
PCB - 153		10 ng/l						
PCB - 156		10 ng/l						
PCB - 157		10 ng/l						
PCB - 167		10 ng/l						
PCB - 169		10 ng/l						
PCB - 170		10 ng/l						
PCB - 180		10 ng/l						
PCB - 187		10 ng/l						
PCB - 189		10 ng/l						
PCB - 195		10 ng/l						
PCB - 206		10 ng/l						
PCB - 209		10 ng/l						